Instructions for cleaning
(Multi-use ONLY)

Step 1: Rinse the RM Electrode Canine gently under tap water immediately after use.

Step 2: Soak for 5 minutes in a mild lab detergent solution (e.g. Liquinox).

Step 3: Rinse under running tap water for five minutes (place in a clean beaker under gently running tap).

Step 4: Rinse with distilled water to avoid water spots and mineral deposits.

Step 5: Air dry.

Step 6: Check to ensure that the electrode ring is in the proper recess within the silicone lens before next use – metal should never touch the eye.